

Automation of Resistance Measurements in a Magnetic Field

Ion DOBROJAN, Dalia POPESCU, Eugeniu SEMENCIUC, Mihai GABOR,
Traian PETRIȘOR

Abstract

In this paper we describe the constituent parts and operation of a highly accurate low resistances automated magnetoresistance measurement system. The system was used to successfully measure the magnetoresistive response of a Co(6nm)/Cu(6nm)/Fe(10nm) spin valve showing a weak effect.

1